

九十四~九十五年度 G6PD 院際品管檢體活性測定的再現性

低活性品管檢體 G6PD 活性測定的再現性

92.12 ~ 95.10

檢驗單位	n ¹	Mean	C.V.(%)
RH-01	7	3.5	14.4
RH-02	7	4.1	5.5
RH-03	7	3.9	10.7
RH-04 ²	7	4.4	59.4
RH-06	7	3.7	9.8
RH-07	7	3.6	6.3
RH-08	7	4.5	28.7
RH-09	7	3.6	17.3
RH-10	7	4.0	16.5
RH-12	7	3.8	4.8
RH-13	7	4.6	24.3
RH-14	7	4.4	29.3
RH-18	7	3.5	8.9
RH-19	7	3.5	6.8
A0189	7	3.7	14.3
A0203	7	3.9	12.9
B001	7	4.0	12.9
G031	7	4.2	20.0
G026	7	3.9	13.2
CL01	7	3.7	14.9
CL02	6	4.1	8.0
CL03	4	4.2	35.9
參考值 ³	7	4.0	6.5

1、送檢批號：93-10、93-11、94-01、94-03、95-01、95-03、95-05

2、RH-04扣除一次異常結果，Mean = 3.4，CV = 24.3%

3、Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital

中活性品管檢體 G6PD 活性測定的再現性

92.12 ~ 95.10

檢驗單位	n ¹	Mean	C.V.(%)
RH-01	4	8.2	5.7
RH-02	4	8.8	6.6
RH-03	4	9.1	11.2
RH-04	4	8.0	25.9
RH-06	4	8.5	11.4
RH-07	4	8.8	7.5
RH-08	4	8.5	21.2
RH-09	4	8.6	11.0
RH-10	4	9.3	7.1
RH-12	4	8.5	5.2
RH-13	4	8.0	3.9
RH-14	4	8.2	7.9
RH-18	4	8.3	14.5
RH-19	4	8.4	9.4
A0189	4	8.4	12.0
A0203	4	8.4	4.4
B001	4	9.1	10.6
G031	4	8.9	11.6
G026	4	9.5	7.5
CL01	4	8.2	9.3
CL02	3	9.8	12.1
CL03	4	9.2	20.9
參考值 ²	4	8.9	3.7

1、送檢批號：92-12、94-01、94-03、94-04

2、Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital

高活性品管檢體 G6PD 活性測定的再現性

92.12 ~ 95.10

檢驗單位	n ¹	Mean	C.V.(%)
RH-01	5	26.4	7.2
RH-02	5	25.4	7.2
RH-03	5	26.8	6.1
RH-04	5	25.4	5.5
RH-06	5	24.7	7.1
RH-07 ²	5	32.8	39.0
RH-08	5	22.8	15.1
RH-09	5	25.3	11.7
RH-10 ³	5	23.2	31.3
RH-12	5	21.2	9.7
RH-13	5	23.8	8.0
RH-14	5	24.0	11.1
RH-18	5	23.2	8.2
RH-19	5	22.3	9.2
A0189	5	25.3	13.1
A0203	5	27.2	3.0
B001	5	24.8	11.3
G031	5	24.6	6.8
G026	5	22.8	7.7
CL01	5	25.4	7.4
CL02	4	24.8	3.6
CL03	5	26.5	17.6
參考值 ⁴	5	27.2	7.6

- 1、送檢批號：93-11、94-01、94-04、94-10、95-03
- 2、RH-07 扣除一次異常結果，Mean = 27.1，CV = 5.4%
- 3、RH-10 扣除一次異常結果，Mean = 26.3，CV = 8.1%
- 4、Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital