

九十六年度 G6PD 院際品管檢體活性測定的再現性

低活性品管檢體 G6PD 活性測定的再現性

96.01 ~ 96.11

| 檢驗單位 | n ¹ | Mean | S.D. | C.V.(%) |
|------------------|----------------|------|------|---------|
| RH-01 | 3 | 5.3 | 0.53 | 10.0% |
| RH-02 | 3 | 4.7 | 0.20 | 4.3% |
| RH-03 | 3 | 4.5 | 0.46 | 10.2% |
| RH-04 | 3 | 4.5 | 0.35 | 7.8% |
| RH-06 | 3 | 5.0 | 0.20 | 4.0% |
| RH-07 | 3 | 4.0 | 0.35 | 8.8% |
| RH-08 | 3 | 5.1 | 0.15 | 2.9% |
| RH-09 | 3 | 4.7 | 0.35 | 7.4% |
| RH-10 | 3 | 5.0 | 0.26 | 5.2% |
| RH-12 | 3 | 5.0 | 0.45 | 9.0% |
| RH-13 | 3 | 4.5 | 0.44 | 9.8% |
| RH-14 | 3 | 4.9 | 0.79 | 16.1% |
| RH-18 | 3 | 4.4 | 0.59 | 13.4% |
| RH-19 | 3 | 5.0 | 0.10 | 2.0% |
| A0189 | 3 | 4.1 | 1.10 | 26.8% |
| A0203 | 3 | 4.6 | 0.53 | 11.5% |
| G031 | 3 | 5.2 | 0.64 | 12.3% |
| G026 | 3 | 5.4 | 0.20 | 3.7% |
| CL01 | 3 | 5.0 | 0.21 | 4.2% |
| CL02 | 3 | 4.4 | 0.55 | 12.5% |
| B001 | 3 | 5.4 | 0.85 | 15.7% |
| 參考值 ² | 3 | 5.2 | 0.15 | 2.9% |

1、 送檢批號：RH2007-03、RH2007-07、RH2007-09

2、 Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital

中活性品管檢體 G6PD 活性測定的再現性

96.01 ~ 96.11

| 檢驗單位 | n ¹ | Mean | S.D. | C.V.(%) |
|------------------|----------------|------|------|---------|
| RH-01 | 3 | 8.4 | 0.17 | 2.0% |
| RH-02 | 3 | 7.7 | 0.35 | 4.5% |
| RH-03 | 3 | 7.4 | 0.36 | 4.9% |
| RH-04 | 3 | 7.5 | 0.46 | 6.1% |
| RH-06 | 3 | 8.0 | 0.35 | 4.4% |
| RH-07 | 3 | 7.9 | 1.02 | 12.9% |
| RH-08 | 3 | 8.3 | 0.72 | 8.7% |
| RH-09 | 3 | 8.1 | 0.56 | 6.9% |
| RH-10 | 3 | 8.2 | 0.36 | 4.4% |
| RH-12 | 3 | 7.5 | 0.84 | 11.2% |
| RH-13 | 3 | 7.7 | 0.58 | 7.5% |
| RH-14 | 3 | 8.1 | 1.10 | 13.6% |
| RH-18 | 3 | 7.2 | 0.61 | 8.5% |
| RH-19 | 3 | 8.2 | 0.42 | 5.1% |
| A0189 | 3 | 7.6 | 1.12 | 14.7% |
| A0203 | 3 | 7.3 | 0.21 | 2.9% |
| G031 | 3 | 9.5 | 0.25 | 2.6% |
| G026 | 3 | 8.9 | 1.10 | 12.4% |
| CL01 | 3 | 8.2 | 0.25 | 3.0% |
| CL02 | 3 | 7.8 | 1.04 | 13.3% |
| B001 | 3 | 9.0 | 0.67 | 7.4% |
| 參考值 ² | 3 | 8.1 | 0.06 | 0.7% |

1、 送檢批號：RH2007-03、RH2007-07、RH2007-09

2、 Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital

高活性品管檢體 G6PD 活性測定的再現性

96.01 ~ 96.11

| 檢驗單位 | n ¹ | Mean | S.D. | C.V.(%) |
|------------------|----------------|------|------|---------|
| RH-01 | 3 | 13.1 | 1.21 | 9.2% |
| RH-02 | 3 | 12.6 | 0.25 | 2.0% |
| RH-03 | 3 | 12.2 | 1.01 | 8.3% |
| RH-04 | 3 | 12.6 | 0.21 | 1.7% |
| RH-06 | 3 | 12.8 | 0.67 | 5.2% |
| RH-07 | 3 | 12.8 | 1.00 | 7.8% |
| RH-08 | 3 | 13.6 | 0.31 | 2.3% |
| RH-09 | 3 | 12.7 | 0.61 | 4.8% |
| RH-10 | 3 | 13.5 | 0.52 | 3.9% |
| RH-12 | 3 | 11.6 | 0.62 | 5.3% |
| RH-13 | 3 | 13.0 | 0.40 | 3.1% |
| RH-14 | 3 | 12.7 | 2.21 | 17.4% |
| RH-18 | 3 | 11.9 | 0.84 | 7.1% |
| RH-19 | 3 | 12.7 | 0.55 | 4.3% |
| A0189 | 3 | 12.9 | 1.89 | 14.7% |
| A0203 | 3 | 11.9 | 0.72 | 6.1% |
| G031 | 3 | 14.2 | 1.27 | 8.9% |
| G026 | 3 | 14.4 | 0.91 | 6.3% |
| CL01 | 3 | 12.4 | 0.36 | 2.9% |
| CL02 | 3 | 13.4 | 1.16 | 8.7% |
| B001 | 3 | 13.2 | 0.70 | 5.3% |
| 參考值 ² | 3 | 12.7 | 0.06 | 0.5% |

1、送檢批號：RH2007-03、RH2007-07、RH2007-09

2、Determined at Clin. Biochem. Res. Lab., Taipei Veterans General Hospital