
## Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - January 2025 -


## I. The statistic results of all laboratories in this month


G6PD	Control N (Lot No.:BJ0922N)	Control D (Lot No.: BJ0922D)
Labs	7	7
Received results number (n)	37	37
Median	14.6 (U/g Hb)	4.7 (U/g Hb)
Mean	14.5 (U/g Hb)	4.8 (U/g Hb)
SD	0.5	0.4
CV	3.4%	8.3%
Range of G6PD	$13.2 \sim 15.5 \text{ (U/g Hb)}$	$4.3 \sim 5.4 \; (\text{U/g Hb})$
Range of Hb	$1.9 \sim 2.6  (g/dL)$	$1.8 \sim 2.3 \; (g/dL)$

<sup>\*</sup>The statistic results are calculated from all labs reported in this month

## II. The distribution of G6PD reported for each lab in this survey









<sup>\*\*</sup> G6PD Method = Medicon reagent kit, 37°C