

Annual Summary of External Quality Assurance  
Survey for Neonatal G6PD Screening Test  
( 2023 )



QAP Center  
Preventive Medicine Foundation  
P.O. Box 624 Taipei Xinwei  
Taipei, Taiwan 10699, R.O.C.

Tel: +886-2-2703-6080  
Fax: +886-2-2703-6070  
Email: [g6pd@pmf.tw](mailto:g6pd@pmf.tw)  
<http://g6pd.qap.tw/en/>


Summary Released : February 07, 2024.



# Annual Summary of External Quality Assurance (EQA) Survey for Neonatal G6PD Screening Test ( 2023 )

## 1. Introduction


Preventive Medicine Foundation Quality Assurance Program Center ( PMF QAP Center ) has been providing “ EQA Program for Neonatal Glucose-6-Phosphate Dehydrogenase ( G6PD ) Screening Test ” using dried blood spot samples for newborn screening laboratories since 1999. This EQA program has been officially accredited by Taiwan Accreditation Foundation ( TAF , a member of ILAC Mutual Recognition Arrangement Signatories ) for conformity to ISO/IEC 17043:2010 since 2017 ( Accreditation No. : P016 ) and was extended the certification on December 30, 2022.



**Figure 1. Certification Accreditation ( Certificate No : P016-221230 )**

## 2. Participants

Fifty- five laboratories ( including 3 reagent manufacturers ) from 13 countries ( AT, BE, CN, DE, FI, GR, IN, MX, PH, TH, TR, TW, and VN ) have participated in the EQA program in 2023.



G6PD Deficiency Worldwide Distribution, WHO working group. Bull WHO 1989;67:601

**Figure 2. Worldwide distribution of participating laboratories.  
( ● Participant, n = 55 )**

The number of participants was the same as last year.




Figure 3. Number of participating laboratories in this program. ( 2009 ~ 2023 ).

### 3. Quality control sample ( QC sample )

- 3.1 Ten QC samples were used in each survey.
- 3.2 The QC samples were prepared by spotting human blood ( no extra G6PD added ) on Whatman 903 filter paper. ( Taiwan IVD Regist. No.: MOHW-MD-(I)-No.004754 )
- 3.3 The homogeneity and stability of QC samples conform to the requirements of international standard ISO/IEC 17043:2010.

### 4. Surveys

- 4.1 There were 3 EQA surveys for screening test performed in 2023. (Table 1)

Table 1. 2023 EQA survey schedule

No.	Survey No	Survey Starting Date *	Reporting Deadline* ( Taiwan )	Reporting Deadline* ( Others )	Survey Result Released*
1	NS2023-01	02/13	02/16	02/23	03/02
2	NS2023-02	05/29	06/01	06/08	06/13
3	NS2023-03	09/11	09/14	09/21	09/28

\* Date: Month/Day

- 4.2 In 2023, 165 sets of QC samples were sent to participants, 164 ( 99.4% ) reports were returned.
- 4.3 Most screening laboratories received the QC samples within 4 ~ 9 days ( Median : 5 days ) after the survey started, which is similar compared to previous years ( 4 ~ 9 days, median : 7 days ).
- 4.4 The report returned time were between 2 and 12 days ( Median : 8 days ) after the survey started, which were compatible with previous years. Most reports ( 95.7% ) were returned within target time ( 10 calendar days after the survey started ).
- 4.5 Only 4.9% ( 8/164 ) of the participants had reported the shipping temperature indicator reached 54.4 °C, which were compatible with previous years. No unsatisfactory result was found due to the shipping temperature.
- 4.6 There are 14.6% ( 24/164 ) participants had reported the shipping humidity indicator reached 30%, which were higher than previous years. No unsatisfactory result was found due to this shipping humidity.
- 4.7 The survey result released between 2 and 6 days ( Median : 6 days ) after reporting deadline, which were compatible to the target time ( 7 working days ).


## 5. Evaluation Criteria reported results

- 5.1 The reported results were evaluated with a consensus decision of the same lot QC samples ( assigned values ) from more than 75% of participants.
- 5.2 The performance evaluation criteria for participant survey report:
- a) Acceptable: all **results** are acceptable ;
  - b) Acceptable with caution: only one **result** is unsatisfactory ;
  - c) Unsatisfactory: more than one **result** are unsatisfactory.

## 6. Overall result of performance evaluation

- 6.1 Overall results of the four EQA surveys for screening test :
- a) 157 ( 95.7% ) reports were “Acceptable” ;
  - b) 4 ( 2.4% ) reports were “Acceptable with caution” ;
  - c) 3 ( 1.8% ) reports were “Unsatisfactory”.

6.2 The unsatisfactory reports rate ( 1.8 % ) was lower than that of 2022 ( 7.9% ) ( Figure 4 ).



**Figure 4. Unsatisfactory rates of the survey reports ( 1999 ~ 2023 ).**

a) There were 3 false negative results ( 1.8% ), and 9 false positive results ( 1.8% ) were reported ( Table 2 ) ;

**Table 2. EQA results of neonatal G6PD screening test at different ranges of G6PD activity in 2023.**

G6PD Activity*	Specimens	Positive (P)	Negative (N)	False P	False N
0.1 – 2.0	54	54	0	0	0
2.1 – 3.2	165	162	3	0	<b>3 (1.8%)</b>
6.0 – 8.0	653	9	644	<b>9 (1.4%)</b>	0
> 8	768	0	768	0	0
<b>Total</b>	<b>1640</b>	<b>225</b>	<b>1415</b>	<b>9 (0.5%)</b>	<b>3 (0.2%)</b>

\* EQA reference laboratory cut off value = 4.4 U/gHb ( using Trinity Biotech 345 reagent at 37°C )

b) The major of unsatisfactory results were caused by the G6PD activity of test samples that close to the cut off value or inappropriate cut off value used by the participants ( Table 2, Table 3 ).

**Table 3.** Reagent kits of G6PD blood spot screening test used by the screening laboratory in 2023 ( Survey : NS2023-03 ).

Reagent kit	No. of laboratory	Cut off value
<b>Quantitative</b>		
Born Safe	1	4 (U/g Hb)
Guangzhou Fenghua	5	2.2, 2.6, 2.7, 2.75, 3.0 (U/g Hb)
Laboratory prepared	2	2.0, 6.2 (U/g Hb)
Labsystems Diagnostics	4	3.0, 3.5, 4.0 (U/g Hb)
PE (ND-1000)	18	2.1, 2.2, 2.5, 2.6 (U/g Hb)
PE GSP Neonatal G6PD (3310-0010)	16	16.0, 17.2, 20.0, 20.5, 22.0, 22.5, 23.7, 24.0, 25.5, 26.0, 29.0 (U/dL)
R&D Diagnostics (OSMMR2000-D)	2	2.5 (U/g Hb)
Zentech	4	2.48, 2.5, 3.64 (U/g Hb)
<b>Qualitative</b>		
R&D Diagnostics (SQMMR500)	2	—
Laboratory prepared	1	—

\* EQA reference laboratory cut off value = 4.4 U/gHb ( using Trinity Biotech 345 reagent at 37°C )

6.3 To evaluate the within laboratory test repeatability of each participant, we compared the performance within a run for each participant who used quantitative test. CVs of “Repeatability of Neonatal G6PD Quantitative Screening Test” were calculated for sample number ≥ 3 from the same QC sample lot. Repeatability (within run precision) can monitor the participant performance in different activity within a run for each laboratory. Repeatability for each survey is summarized in Table 4.

- a) The median of CVs for within laboratory repeatability were between 3.4% and 4.7%. ( Table 4 ) ;
- b) The range of CVs for within laboratory repeatability was between 0% and 31.1% ( Table 4 ) ;
- c) Most screening laboratories present a good within laboratory repeatability ( CV < 10% ) using neonatal G6PD quantitative screening test.

**Table 4.** Repeatability of Neonatal G6PD Quantitative Screening Test.

Reagent code <sup>1</sup>	NS2023-03 <sup>2</sup>		NS2023-01 <sup>2</sup>		NS2023-02 <sup>2</sup>		NS2023-02 <sup>2</sup>		NS2023-03 <sup>2</sup>	
	2.3 U/gHb <sup>3,4</sup>		6.0 U/gHb <sup>3,4</sup>		6.4 U/gHb <sup>3,4</sup>		11.3 U/gHb <sup>3,4</sup>		15.3 U/gHb <sup>3,4</sup>	
	n <sup>5</sup>	CVs (%) <sup>3,6</sup>	n <sup>5</sup>	CVs (%) <sup>3,6</sup>	n <sup>5</sup>	CVs (%) <sup>3,6</sup>	n <sup>5</sup>	CVs (%) <sup>3,6</sup>	n <sup>5</sup>	CVs (%) <sup>3,6</sup>
15	18	4.4% (0 - 10.0)	17	4.7 (2.7 - 31.1)	17	5.0 (1.3 - 20.0)	17	5.1 (0.8 - 14.1)	18	3.5 (2.3 - 15.9)
19	16	5.3 (1.8 - 8.7)	16	6.4 (1.8 - 16.0)	17	3.2 (1.0 - 9.5)	17	3.7 (1.8 - 15.1)	16	6.0 (2.7 - 20.1)
18	5	4.4 (1.2 - 7.6)	4	3.1 (2.3-4.1)	5	2.3 (1.9 - 7.6)	5	2.8 (1.6 - 2.8)	5	2.8 (2.0 - 3.3)
20	4	7.0 (2.2 - 10.5)	4	4.0 (2.7 - 10.4)	3	7.1 (2.2 - 10.0)	3	5.2 (4.3 - 5.6)	4	9.6 (2.2 - 19.1)
21	4	6.8 (2.5 - 7.5)	3	5.8 (2.4 - 9.5)	4	2.5 (1.6 - 3.0)	4	4.3 (2.6 - 6.0)	4	3.8 (1.6 - 5.9)
3	2	(1.3 , 2.9)	2	(3.1 , 6.6)	2	(0 , 8.1)	2	(6.8 , 7.3)	2	(3.7 , 7.3)
99	2	(7.0 , 7.1)	3	3.7 (1.5 - 19.0)	3	4.2 (1.3 - 4.4)	3	3.9 (0.7 - 8.9)	2	(6.4 , 15.3)
22	1	10.0	1	4.9	1	20.0	1	3.6	1	14.5
Total	52	4.7 (0 - 10.5)	50	4.7 (1.5 - 31.1)	52	3.4 (0 - 20.0)	52	4.1 (0.7 - 15.1)	52	4.5 (1.6 - 20.1)

Note : 1. Reagent code : <https://g6pd.qap.tw/NSdata.php?BatchNo=NS2023-03>

2. Survey number

3. CVs were calculated for sample ≥3 from the same sample lot for each participant.

4. The reference values were determined by using Trinity Biotech 345-UV reagent at the QC reference Laboratory.

5. n : number of participants

6. Median (Range of CVs.)

6.4 All the results of EQA surveys for screening test in 2023 were posted on website < <https://g6pd.qap.tw/112nsrep-eng.htm> >.

### 7. Conclusion of customer satisfaction survey

In 2023, 55 customer satisfaction surveys were sent to participants, 43 ( 78% ) questionnaires were returned. Among the returned questionnaires, 77% of the participants rate “Excellent” performance and 23% of the participants rate “Great” performance in overall satisfaction.

